MECHANICAL DESIGN OF A DIAMOND CRYSTAL HARD X-RAY SELF-SEEDING MONOCHROMATOR FOR PAL-XFEL*

D. Shu[†], Y. Shvyd'ko, S. Kearney, J. Anton, K-J. Kim, Argonne National Laboratory,

Argonne, IL, USA

B.G. Oh, C-K. Min, S.Y. Rah, H-S. Kang, Pohang Accelerator Laboratory. Pohang, Gyeongbuk, Republic of Korea

Abstract

author(s), title of the work, publisher, and DOI As a part of the Argonne Strategic Partnership Project (SPP) 85H21, a collaboration between Advanced Photon Source (APS), Argonne National Laboratory (ANL) and 2 Pohang Accelerator Laboratory (PAL), we have designed, 5 constructed, and tested a thin-diamond-crystal monochromator for the PAL X-ray Free-Electron-Laser (PAL-XFEL) hard x-ray self-seeding project [1]. The mechanical design of the PAL-XFEL diamond crystal hard x-ray selfg seeding monochromator is based on the ArS design of a thin-diamond-crystal monochromator for the LCLS hard xray self-seeding project [2,3] with enhanced diamond crysray self-seeding project [2,3] with enhanced diamond crys-tal holder for two thin-diamond crystals with thicknesses ₹ of 30 microns and 100 microns [4]. The customized high auality thin-diamond-crystals and special graphite holder quality thin-diamond-crystals and special graphite holder " were provided by the Technological Institute for Superbard and Novel Carbon Materials of Russia (TISNCM) [5], 5 and tested at the APS [4]. An in-vacuum multi-axis precision positioning mechanism is designed to manipulate the duo-thin- diamond-crystals holder with resolutions and sta-¹⁷ bilities required by the hard x-ray self-seeding physics. Dechanical specifications, designs, and preliminary test results of the diamond monochromator are presented in this paper.

INTRODUCTION

3.0 licence (© 2019). This short To longitudinally improve the coherent quality of the x-ray radiation produced by the hard X-ray Free-Electron-Laser (XFEL), and to reduce the level of the spik- \gtrsim iness in their spectrum and temporal structures, in 2001 Saldin and et al. proposed the initial idea of self-seeding based on the utilization of Bragg diffraction (BD) from a four-bounce diamond crystal monochromator as a bandpass filter [6]. About ten years later, Geloni and et al. proposed a transmission self-seeding using forward Bragg deflection (FBD) from a single diamond crystal with simple alignment requirements [7]. er

The first hard X-ray self-seeding with forward Bragg deflection (FBD) from a single diamond crystal was demonstrated at the Linac Coherent Light Source (LCLS) in 2012 2[2]. Based on the experiences gained from the first LCLS ⇒hard x-ray self-seeding (HXRSS) monochromator design [3], a thin-diamond-crystal monochromator for the PAL-XFEL hard x-ray self-seeding project has been designed, s constructed, and commissioned as a part of the Argonne

Strategic Partnership Project (SPP) 85H21, a collaboration between APS, ANL and PAL.



Figure 1: 3D model of the PAL-XFEL self-seeding monochromator integrated on the undulator girder system.

The monochromator was manufactured by a Korean company (VACTRON [8]) based on the ANL/PAL collaborated engineering design, and was installed at the PAL-XFEL in February, 2018. Manufactured by the TISNCM and tested at the APS, the enhanced monochromator diamond crystal holder accommodates two customized high quality thin-diamond-crystals with thickness of 30 µm and 100 µm in the [10] and [11] orientations to satisfy various optimization conditions [4, 9]. The PAL-XFEL self-seeding monochromator has been successfully commissioned without any design flaws soon after its installation [1].

Mechanical specifications, designs, and preliminary test results of the diamond crystal monochromator for the PAL-XFEL are presented in this paper.



Figure 2: Photograph of the PAL-XFEL self-seeding monochromator integrated on undulator girder system.

COMPACT UHV ENCLOSURE

A compact vacuum enclosure was designed for the PAL-XFEL hard x-ray self-seeding monochromator with ultrahigh-vacuum (UHV) compatibility as required by PAL-XFEL vacuum system. As shown in Fig. 1, the hard x-ray self-seeding monochromator is integrated on one of the

MC6: Beam Instrumentation, Controls, Feedback and Operational Aspects **T26** Photon Beam Lines and Components

from 1 * Work supported by the U.S. Department of Energy, Office of Science, under Contract No. DE-AC02-06CH11357.

Content † shu@aps.anl.gov

10th Int. Partile Accelerator Conf. ISBN: 978-3-95450-208-0

PAL-XFEL undulator girder system with four dipole chicane magnets and various diagnostic tools including: stripline beam position monitor, wire scanner, in-line screen monitor, surveillance camera and et al. The selfseeding monochromator system fits into one original undulator girder as a modular design so that it can be replaced with any undulator and relocated in the future. Figure 2 shows a photograph of the PAL-XFEL self-seeding monochromator integrated on the undulator girder system.

As shown in Fig. 3, the monochromator vacuum enclosure hosts a 12" flange for a 4-axis diamond positioning system, a beryllium window for x-ray diagnostics, a 25 l/s ion pump, two viewports, and a pair of UHV bellows connected to the PAL-XFEL vacuum system. The distance between the flanges for bellows is limited to less than 212 mm. The monochromator vacuum enclosure is mounted to the girder system through a base plate with alignment mechanisms.



Figure 3: 3D model of the PAL-XFEL self-seeding monochromator UHV enclosure.

PRECISION DIAMOND CRYSTAL MANIPULATION IN UHV

Similar to the LCLS self-seeding monochromator design, a 4-axis precision stage system for diamond crystal manipulation, which includes a precision rotary stage for diamond crystal pitch motion control and a set of tip-tilt and linear stages is mounted on a 12" base flange as shown in Fig. 4. The 12" base flange is equipped with electric feedthroughs for various in-vacuum motors and optical encoders closed-loop controlled from controllers outside vacuum. Precision mounting holes are also prepared for diamond-crystal position survey and alignment from outside vacuum.

Pitch Rotary Stage

The self-seeding monochromator pitch stage controls the Bragg angle of the diamond crystal. It is a customized PITM PRS-110 UHV-compatible stepping-motor-driven rotary stage with a 0.0001-degree resolution optical angular encoder [10]. The customized pitch stage is specified to have a better than 0.0002-degree unidirectional repeatability with closed-loop control. To protect the optical encoder and electronics components in the stage, molybdenum radiation shielding plates are applied to the rotary stage as shown in Figure 4.

T26 Photon Beam Lines and Components



Figure 4: 3D model of the PAL-XFEL self-seeding monochromator 4-axis precision stage system.

Tip-tilt and Linear Stages

As shown in Fig. 4, the PAL-XFEL self-seeding monochromator tip-tilt stage provides a roll adjustment for the diamond crystal alignment. Similar to the LCLS self-seeding monochromator design, the 4-bar flexural bearing structure is modified from an original APS design for a compact multi-dimensional alignment apparatus developed for multilayer Laue lenses (MLLs) with nanometerscale 2-D focusing [2]. The flexure tip-tilt stage is operated by a SmarActTM PZT-driven linear stage [11] to provide a precise angular positioning around the Y_{roll}-axis, which is rotating with the pitch rotary stage, and agrees with the Yaxis while the pitch rotary angle is at the 90-degree position.

Table 1: Design Specifications for The PAL-XFELHXRSS Monochromator 4-axis Precision Stage System

Stage Design Specifications	
Crystal pitch angle operation range (deg)	32 - 95
Crystal pitch angle limit switch range (deg)	31 - 96
Crystal pitch angle hard limit range (deg)	29 - 97
Pitch angle stability (rms mrad)	< 0.005
Crystal roll angle control range (deg)	-5.5 - +4.8
Crystal roll angle stability (rms mrad)	< 0.010
X position control range (mm)	-2.8 - +4.8
Yroll position control range (mm)	-12.6 - +2.1
X and Yroll position stability (rms mm)	< 0.006
Crystal extraction position (approx. mm)	-11
Vacuum	UHV
F-F distance (mm)	212

The X and Y_{roll} linear stages, mounted on top of the flexure tip-tilt stage, provide the linear motion required by the PAL-XFEL self-seeding monochromator system. During the self-seeding test, the X and Y_{roll} linear stages align the diamond crystal to the test position. With safety interlock

Content

10th Int. Partile Accelerator Conf. ISBN: 978-3-95450-208-0

control, the X and Y_{roll} linear stages can retract the diamond ¹/₂ crystal holder to a safe position in the PAL-XFEL normal ¹/₂ operation condition. Table 1 summarizes the design speci-¹/₂ fications for the monochromator pitch rotary stage system. fications for the monochromator pitch rotary stage system. Figure 5 shows a photograph of the PAL-XFEL self-seeding monochromator 4-axis precision stage system.



Figure 5: Photograph of the PAL-XFEL self-seeding monochromator 4-axis precision stage system with a dummy of diamond crystals holder for stage test.

THIN DIAMOND CRYSTAL PLATE HANDLING

Since the diamond crystal holder for PAL-XFEL selfseeding monochromator is close to the electron beam in the linear accelerator, it is best that the diamond holder be Emade of low-Z materials, such as highly ordered pyrolytic graphite (HOPG) to meet radiation safety requirements. S The new design of diamond crystal holder for PAL-XFEL R self-seeding monochromator kept the same toothbrush-Shaped HOPG mechanical design of the LCLS self-seeding

- Shaped HOPG mechanical design of the LCLS self-seeding monochromator with several important design improvements as shown in Fig. 6:
 The new holder is designed to host two thin-film-diamonds with thickness of 30 µm and 100 µm in the [10] and [11] orientations to satisfy various optimization conditions at the PAL-XFEL.
 The new holder includes a HOPG holder base and a CVD diamond mounting base to ensure a good heat transfer between the diamond crystal and the heat sink.
 CVD diamond post and thin graphite springs are used to hold the thin diamond crystals and provide a gentle clamping force between the thin diamond crystal and CVD diamond mounting base to ensure a good thermal contact with minimized strain added on the thin diamond crystal.

THPGW089 MC6: Ream Instru



Figure 6: 3D model of the PAL-XFEL self-seeding monochromator thin diamond crystal holder.

SUMMARY

We presented mechanical specifications and designs of the thin-diamond-crystal monochromator for the PAL-XFEL hard x-ray self-seeding project. In April, 2018, the PAL-XFEL commissioning team led by Heung-Sik Kang demonstrated the first seeded FEL signal (at ~8.4keV) as shown in Figure 7. The results showed spectra amplitude of the self-seeding FEL with four times higher than SASE at Angstrom wavelengths. The bandwidth of self-seeding FEL is as small as 0.5 eV. Further testing and development of the PAL-XFEL hard x-ray self-seeding project for user operation are forthcoming.



Figure 7: Spectra of the preliminary self-seeding test with PAL-XFEL hard x-ray self-seeding monochromator.

ACKNOWLEDGMENT

The authors would like to acknowledge the management support from Argonne National Laboratory and Pohang Accelerator Laboratory. Work supported by the U.S. Department of Energy, Office of Science, under Contract No. DE-AC02-06CH11357.

REFERENCES

- [1] Chang-Ki Min *et al.*, "Hard X-ray self-seeding commissioning at PAL-XFEL", submitted for publication.
- [2] D. Shu *et al., J. Phys.: Conf. Ser.*, vol. 425, no. 052004, 2013 doi:10.1088/1742-6596/425/5/052004
- [3] J. Amann et al., Nat. Photon. 6, pp. 693–698, 2012.
- [4] Yu. Shvyd'ko et al., "Diamond Double-Crystal System for a Forward Bragg Diffraction X-Ray Monochromator of the Self-Seeded PAL XFEL", in Proc. 38th Int. Free Electron Laser Conf. (FEL'17), Santa Fe, NM, USA, Aug. 2017, pp. 29-33, doi:10.18429/JACOW-FEL2017-MOP001
- [5] S. Polyakov *et. al., Diamond and Related Materials*, vol. 20, no. 726, 2011.
- [6] E. L. Saldin, E.A. Schneidmiller, Y. V. Shvyd'ko and M.V. Yurkov, *Nucl. Instrum. Methods Phys. Res. A*, vol. 475, pp. 357–362, 2001.
- [7] G. Geloni, V. Kocharyan, and E. Saldin, J. Mod. Opt., vol. 58, no. 1391, 2011.
- [8] https://www.vactron.co.kr/en/
- [9] X. Yang and Y.V. Shvyd'ko, *Phys. Rev. ST Accel. Beams*, vol. 16, no. 120701, 2013.
- [10] https://www.pi-usa/en/
- [11] https://www.smaract.com
- [12] S. Terentyev, V. Blank, T. Kolodziej, and Y. Shvyd'ko. "Curved diamond-crystal spectrographs for x-ray free-electron laser non-invasive diagnostics", *Rev. Sci. Instrum.*, vol. 87, no. 125117, 2016.